

**Air Entrainment Meter for  
Mortar**



**CEMENT**

## Air Entrainment Meter for Mortar



### PRODUCT MODEL

M2410 Air Entrainment Meter for Mortar

### PRODUCT STANDARDS

Standards EN 413-2, 459-2, 1015-7 | DIN 18 555

### INFORMATION

Manufacturer	TESTMAK INS.LAB.MAK.SAN.VE TIC. PAZ. ITH. IHR. LTD. STI
Country of Origin	TURKEY
Product name	Air Entrainment Meter for Mortar

## Air Entrainment Meter for Mortar

### DESCRIPTION

The Testmak M2410 model Air Entrainment Meter for Mortar is used for determining the air content of cement paste, cement mortar and lime mortar. The air entrainment meter is manufactured from cast aluminum, the upper part and the lower test pot are held together with an air-tight seal which are easily adjusted by using the two spring clamps. The pressure gauge is installed in the head of the meter and the scale works in the 0-20 volumetric % range. The air is compressed with a hand pump installed in the system and the smart configuration of the test and correction buttons enables fast and simple testing.

### TECHNICAL SPECIFICATIONS

P. Code	M2410
Dimensions(mm)	200x200x320 mm
Weight (kg)	3.5 kg





**TM TEST MAK**  
**MATERIAL TEST EQUIPMENTS**

Telephone : +90 312 395 36 42  
Fax : +90 312 395 36 01

Website : [www.testmak.com](http://www.testmak.com)  
E-Mail : [info@testmak.com](mailto:info@testmak.com)

Koca Sinan Industrial Sites 1183 Street No:  
40 06370 OSTIM / ANKARA / TURKEY