



BOHME ABRASION

TESTER

AG1160



BOHME ABRASION TESTER





PRODUCT MODEL

AG1160	Böhme Abrasion Tester, 220-240 V 50/60 Hz	
AG1160/110	Böhme Abrasion Tester, 110 V 60 Hz	
AG1162	Abrasive white corundum sand 80 grade. Pack of 25 kg	

PRODUCT STANDARDS

Standards	EN 1338:2004 EN 1339 EN 1340 EN 13892-3 EN 14157 DIN 52108

INFORMATION

Manufacturer	TESTMACLAB LABORATUVAR TEST CİHAZLARI PAZ. VE DIŞ TİC. LTD. ŞTİ	
Country of Origin	TÜRKİYE	
Product name	Böhme Abrasion Tester	



Telephone : +90 312 395 44 57

Fax : +90 312 395 36 42

Website : www.testmak.com E-Mail : info@testmak.com Ağaç Metal Sanayi Sitesi 1354 Cadde 1436 Sokak No:16 Ivogsan Yenimahalle - ANKARA / TÜRKİYE

BOHME ABRASION TESTER



DESCRIPTION

Böhme Abrasion Tester is used for determining the abrasion resistance of paving stones, concrete slabs, natural rocks, natural stone slabs.

The Böhme machine adjustable charger used produce a force of 294N ± 3N on a specimen. The machine is equipped with a display control panel that allows the setting of speed and number of cycles. Once the machine is started, it will automatically stop at the end of the set number of cycles programmed on the controls panel.

Abrasive sand must be ordered separately.

The AG1160 Böhme abrasion tester is essentially composed of:

- A grinding wheel approximately 750 mm in diameter, a removable test mass of 30 kg.
- Clamping device for the sample.
- · Digital control panel

OPERATION

- 1- 20g of standard abrasive sand is placed on the test mark.
- 2- The sample is fixed to the clamping device on the abrasive disc, with standard abrasive sand sprinkled on the trace.



3- The position of the cleaning brushes is adjusted.



Telephone: +90 312 395 44 57 Fax: +90 312 395 36 42 Website: www.testmak.com
E-Mail: info@testmak.com

Ağaç Metal Sanayi Sitesi 1354 Cadde 1436 Sokak No:16 Ivogsan Yenimahalle - ANKARA / TÜRKİYE

BOHME ABRASION TESTER



4- The RPM value is entered from the digital panel and the start button is pressed.



- 5- The disc is rotated at 30 ± 1 rpm, applying an abrasion force of (294 ± 3) N to the sample.
- 6- The decrease in sample volume is calculated.

TECHNICAL SPECIFICATIONS

Dimensions	1400x900x1300 mm
Weight	320 kg
Power	220-240 V 50/60 Hz 750 W



Telephone: +90 312 395 44 57 Fax: +90 312 395 36 42 Website : www.testmak.com E-Mail : info@testmak.com



THANK YOU

FOR CHOOSE US

Dear business partners, thank you very much for believing in us and recommending our products to your customers; We sincerely believe that our company will grow even more thanks to you, our esteemed business partners.

You can reach us than our phones or e mail address 24 hours a day.

CONTACT US



Ağaç Metal Sanayi Sitesi 1354 Cadde 1436 Sokak No:16 Ivogsan Yenimahalle -ANKARA / TÜRKİYE



info@testmak.com marketing@testmak.com



TÜRKİYE / HEAD Office : +90 312 395 44 57 UKRAINE Office : +380 63 741 29 20 SOUTH AMERICA / Cuba Office : +53 5 073 96 12



www.testmak.com