



LENGTH ELENGATION GAUGE

AG1050





LENGTH ELENGATION GAUGE

PRODUCT MODEL

AG1050 Length Elongation Gauge

PRODUCT STANDARDS

Standards BS 812-105.2



INFORMATION

Manufacturer	TESTMACLAB LABORATUVAR TEST CİHAZLARI PAZ. VE DIŞ TİC. LTD. ŞTİ
Country of Origin	TÜRKİYE
Product name	Length Elongation Gauge

DESCRIPTION

Length Gauge is used for determining the elongation index of aggregates. The particle is elongated when its length is more than 1.8 o the midsize of the sieve fraction. The aggregate to be classified is separated into six sieve fractions from 6.3 to 50 mm and each fraction is examined separately.

TECHNICAL SPECIFICATIONS

Dimensions	370x70x70 mm
Weight	1 kg



Telephone : +90 312 395 44 57

Fax : +90 312 395 36 42

Website: www.testmak.com
E-Mail: info@testmak.com

Ağaç Metal Sanayi Sitesi 1354 Cadde 1436 Sokak No:16 Ivogsan Yenimahalle - ANKARA / TÜRKİYE



THANK YOU

FOR CHOOSE US

Dear business partners, thank you very much for believing in us and recommending our products to your customers; We sincerely believe that our company will grow even more thanks to you, our esteemed business partners.

You can reach us than our phones or e mail address 24 hours a day.

CONTACT US



Ağaç Metal Sanayi Sitesi 1354 Cadde 1436 Sokak No:16 Ivogsan Yenimahalle -ANKARA / TÜRKİYE



info@testmak.com marketing@testmak.com



TÜRKİYE / HEAD Office : +90 312 395 44 57 UKRAINE Office : +380 63 741 29 20 SOUTH AMERICA / Cuba Office : +53 5 073 96 12



www.testmak.com