



Vicat Test Apparatus (Manual)

CEMENT

Vicat Test Apparatus (Manual)



PRODUCT MODEL

TMM-2450E	Vicat Test Set, EN
TMM-2450A	Vicat Test Set, ASTM
TMM-2451	Vicat Frame
TMM-2452	Vicat Mould EN Base ID 80 mm, Top ID 70 mm, Height 40 mm
TMM-2453	Vicat Mould ASTM Base ID 70 mm, Top ID 60 mm, Height 40 mm
TMM-2454	Initial Vicat Needle, EN 1.13 mm diameter
TMM-2455	Vicat Needle, ASTM 1 mm diameter
TMM-2456	Final Vicat Needle, EN 1.13 mm diameter
TMM-2457	Consistency Plunger 10 mm diameter
TMM-2458	Supporting Glass Plate
TMM-2459	Additiniol Weight, 700 gr, EN

STANDARDS

ASTM C187, C191 | AASHTO T131 | EN 196-3:2005 | EN 408-2

INFORMATION

Manufacturer : TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ
Country of Origin : TURKEY
Product Name : Vicat Apparatus (Manual)
Code of Product : TMM-2450E, TMM-2450A

Vicat Test Apparatus (Manual)

DESCRIPTION

The Vicat Test Apparatus is used to determine the amount of water required to produce a cement paste of standard consistency and the setting time. The frame is supplied complete with a 300 g consistency plunger (dia. 10 mm). The measurement of the movement is given by an indicator which moves along a scale graduated in millimetres. The apparatus is supplied complete with glass plate and EN Needle and mould.

The Vicat Test Set is supplied complete with;

- Vicat Mould
- Initial and final Needles (for EN)
- Vicat Needle (for ASTM)
- Consistency Plunger
- Supporting Plate
- Glass Thermometer 110°C
- Transfer Dish (for EN)

TECHNICAL SPECIFICATIONS

Dimensions	260x250x450 mm
Weight	4 kg





Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY