

**MANUAL LIQUID LIMIT
DEVICE (CASAGRANDE)**



SOIL

Manual Liquid Limit Device (Casagrande)



PRODUCT MODEL

- TMS-4250 | Manual Liquid Limit Device (Casagrande), BS
- TMS-4252 | Manual Liquid Limit Device (Casagrande), ASTM - AASHTO
- TMS-4253 | Metal Grooving Tool and Gauge Block, ASTM
- TMS-4254 | Plastic Grooving Tool, ASTM
- TMS-4255 | Plastic Grooving Tool, BS
- TMS-4256 | Brass Grooving Tool, AASHTO
- TMS-4257 | Spare Brass Cup, for all Models

STANDARDS

ASTM D4313 | AASHTO T89 | BS 1377:2

INFORMATION

Manufacturer : TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ

Country of Origin : TURKEY

Product Name: Manual Liquid Limit Device (Casagrande)

Code of Product: TMS-4250, 4252

HS CODE: 90248099

Manual Liquid Limit Device (Casagrande)

DESCRIPTION

The TMS-4450 and TMS-4452 Manual Liquid Limit Device (Casagrande) are used to determine the moisture content at which clay soils pass from plastic to liquid state.

The Devices consist of an adjustable crank and cam mechanism, a blow counter and a removable brass cup fitted on the base. Different models with the same shape but with different base and cup weights are available according to the required specifications.

TMS-4250 model is supplied with a BS type plastic grooving tool. TMS-4252 model are supplied with a ASTM type plastic grooving tool. Other types of grooving tools should be ordered separately.

TECHNICAL SPECIFICATIONS

Technical Specifications	
Dimensions	200x220x150 mm
Weight (approx.)	2 kg



TESTMAK
MATERIAL TEST EQUIPMENTS

Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY