



## Vicat Test Set

**CEMENT**

## Vicat Test Set



### PRODUCT MODEL

M2450/E	Vicat Test Set, EN
M2450/A	Vicat Test Set, ASTM
M2450/01	Vicat Frame
M2450/02	Consistency Plunger 10 mm diameter
M2450/03	Supporting Glass Plate
M2450/E-04	Vicat Mould EN Base ID 80 mm, Top ID 70 mm, Height 40 mm
M2450/A-04	Vicat Mould ASTM Base ID 70 mm, Top ID 60 mm, Height 40 mm
M2450/E-05	Initial Vicat Needle, EN 1.13 mm diameter
M2450/A-05	Vicat Needle, ASTM 1 mm diameter
M2450/E-06	Final Vicat Needle, EN 1.13 mm diameter
M2450/E-07	Additiniol Weight, 700 gr, EN

### PRODUCT STANDARDS

Standards ASTM C187, C191 | AASHTO T131 | EN 196-3:2005 | EN 408-2

### INFORMATION

Manufacturer	TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ
Country of Origin	TURKEY
Product Name	Vicat Test Set

## Vicat Test Set

### DESCRIPTION

The Vicat Test Apparatus is used to determine the amount of water required to produce a cement paste of standard consistency and the setting time. The frame is supplied complete with a 300 g consistency plunger (dia. 10 mm). The measurement of the movement is given by an indicator which moves along a scale graduated in millimetres. The apparatus is supplied complete with glass plate and EN Needle and mould.

**The Vicat Test Set is supplied complete with;**

- Vicat Mould
- Initial and final Needles (for EN )
- Vicat Needle (for ASTM)
- Consistency Plunger
- Supporting Plate
- Glass Thermometer 110°C
- Transfer Dish (for EN)

### TECHNICAL SPECIFICATIONS

Dimensions(mm)	260x250x450 mm
Weight (kg)	4 kg





**TM TEST MAK**  
**MATERIAL TEST EQUIPMENTS**

Telephone : +90 312 395 36 42  
Fax : +90 312 395 36 01

Website : [www.testmak.com](http://www.testmak.com)  
E-Mail : [info@testmak.com](mailto:info@testmak.com)

Koca Sinan Industrial Sites 1183 Street No:  
40 06370 OSTIM / ANKARA / TURKEY