

**Compaction Factor
Apparatus**



CONCRETE

Compaction Factor Apparatus



PRODUCT MODEL

C3280 BCompaction Factor Apparatus

STANDARDS

Standards BS 1881-103, 5075 | EN 12390-3

INFORMATION

Manufacturer	TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ
Country of Origin	TURKEY
Product Name	Compaction Factor Apparatus

Compaction Factor Apparatus

DESCRIPTION

The C3280 Compaction Factor Apparatus is designed to undertake a more precise and sensitive test procedure than the simple slump test. The compacting factor is the ratio between the weight of the partially compacted concrete and the weight of the fully compacted concrete. The apparatus consists from two conical hoppers mounted on a cylinder. The hoppers and the mould are mounted onto a rigid steel frame and are easily removable for cleaning.

TECHNICAL SPECIFICATIONS

Product Code	C3280
Dimensions	320x420x1300 mm
Weight	45 kg





TM TEST MAK
MATERIAL TEST EQUIPMENTS

Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY