



FLAKINESS THICKNESS GAUGE

AGGREGATE

Flakiness Thickness Gauge



PRODUCT MODEL

| | |
|--------|---------------------------|
| AG1040 | Flakiness Thickness Gauge |
|--------|---------------------------|

PRODUCT STANDARDS

| | |
|-----------|--------------|
| Standards | BS 812-105.1 |
|-----------|--------------|

INFORMATION

| | |
|-------------------|---|
| Manufacturer | TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ |
| Country of Origin | TURKEY |
| Product name | Flakiness Thickness Gauge |

DESCRIPTION

Thickness Gauge is used to determine if the aggregate particles are to be considered as flaky, i.e. their thickness is less than 0.6 of their nominal size.

The aggregate to be classified is separated into seven sieve fractions from 6,3 to 63 mm.

Slot sizes are 4.9 x 30 mm, 7.2 x 40 mm, 10.2 x 50 mm, 14.4 x 60 mm, 19.7 x 80 mm, 26.3 x 90 mm and 33.9 x 100 mm.

TECHNICAL SPECIFICATIONS

| | |
|------------|---------------|
| Dimensions | 310x130x10 mm |
| Weight | 0,6 kg |



Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY