



Vicat Test Set

CEMENT

Vicat Test Set



PRODUCT MODEL

M2450/E	Vicat Test Set, EN
M2450/A	Vicat Test Set, ASTM
M2450/01	Vicat Frame
M2450/02	Consistency Plunger 10 mm diameter
M2450/03	Supporting Glass Plate
M2450/E-04	Vicat Mould EN Base ID 80 mm, Top ID 70 mm, Height 40 mm
M2450/A-04	Vicat Mould ASTM Base ID 70 mm, Top ID 60 mm, Height 40 mm
M2450/E-05	Initial Vicat Needle, EN 1.13 mm diameter
M2450/A-05	Vicat Needle, ASTM 1 mm diameter
M2450/E-06	Final Vicat Needle, EN 1.13 mm diameter
M2450/E-07	Additiniol Weight, 700 gr, EN

PRODUCT STANDARDS

Standards ASTM C187, C191 | AASHTO T131 | EN 196-3:2005 | EN 408-2

INFORMATION

Manufacturer	TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ
Country of Origin	TURKEY
Product Name	Vicat Test Set

Vicat Test Set

DESCRIPTION

The Vicat Test Apparatus is used to determine the amount of water required to produce a cement paste of standard consistency and the setting time. The frame is supplied complete with a 300 g consistency plunger (dia. 10 mm). The measurement of the movement is given by an indicator which moves along a scale graduated in millimetres. The apparatus is supplied complete with glass plate and EN Needle and mould.

The Vicat Test Set is supplied complete with;

- Vicat Mould
- Initial and final Needles (for EN)
- Vicat Needle (for ASTM)
- Consistency Plunger
- Supporting Plate
- Glass Thermometer 110°C
- Transfer Dish (for EN)

TECHNICAL SPECIFICATIONS

Dimensions(mm)	260x250x450 mm
Weight (kg)	4 kg





TM TEST MAK
MATERIAL TEST EQUIPMENTS

Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY