



Blaine Air Permeability Apparatus

CEMENT

Blaine Air Permeability Apparatus



PRODUCT MODEL

M2470	Blaine Air Permeability Test Apparatus Set
M2470/01	Wood Test Stand
M2470/02	Manometer U Shaped for Blaine Apparatus
M2470/03	Manometer Liquid 250 ml
M2470/04	Cell with Perforated Disc and Plunger
M2470/05	Filter Paper 100 pcs
M2470/06	Reference Cement 5 g
M2470/07	Plastic Funnel 40x65 mm

PRODUCT STANDARDS

Standards	EN 196-6 ASTM C204 AASHTO T153
-----------	------------------------------------

INFORMATION

Manufacturer	TESTMAK INS.LAB.MAK.SAN.VE TİC. PAZ. İTH. İHR. LTD. STİ
Country of Origin	TURKEY
Product Name	Blaine Air Permeability Test Apparatus Set

Blaine Air Permeability Apparatus

DESCRIPTION

The M2470 Blaine Air Permeability Apparatus is used to determine the fineness of Portland cement, limes and similar powders expressed in terms of their specific surface. Reference Cement should be ordered separately.

The M2470 Blaine Air Permeability Test Apparatus Set is supplied complete with;

- Wood Test Stand,
- Rubber Aspirator,
- A Glass Connection Parts,
- Cell with Perforated Disc and Plunger,
- U Manometer Tube,
- Manometer Liquid
- Plastic funnel
- Filter paper (100 pcs of pack)

TECHNICAL SPECIFICATIONS

P. Code	M2470
Dimensions(mm)	230x180x470 mm
Weight (kg)	6 kg





TESTMAK
MATERIAL TEST EQUIPMENTS

Telephone : +90 312 395 36 42
Fax : +90 312 395 36 01

Website : www.testmak.com
E-Mail : info@testmak.com

Koca Sinan Industrial Sites 1183 Street No:
40 06370 OSTIM / ANKARA / TURKEY